

File :Q:\GCMS-06 Minimal\DATA\032120\13- .D
Operator : Organic QC Lab
Acquired : 23 Mar 2020 13:42 using AcqMethod CHICK_2017_13MIN.M
Instrument : GCMS 6
Sample Name: CLP-LC-IS-10X
Misc Info : Laboratory Control Sample in Methanol
Vial Number: 125



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

