



File :K:\QC Data\GCMS-04 Minimal\DATA\031021\25 RM.D
Operator : ORGANIC QC DEPT.
Acquired : 10 Mar 2021 20:51 using AcqMethod VOC-200SPLIT HOT280.M
Instrument : GCMS4
Sample Name: CLP-PI-5X
Misc Info :
Vial Number: 62

Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

