



Acq. Operator : SYSTEM Seq. Line : 9
Acq. Instrument : LCMS-1 Location : P1-E5
Injection Date : 7/1/2021 3:31:19 PM Inj : 1
Inj Volume : 1.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.500 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 5/21/2021 4:03:06 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 7/2/2021 11:19:35 AM by SYSTEM
(modified after loading)

Method Info : Operator: Organic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
MS Detector. No UV Reference.

Sample Info : Maryland Cannabis Pesticide Standard @100ug/mL in AcCN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

