


File :Q:\GCMS-06 Minimal\DATA\070121\14- KM.D
Operator : Organic QC Lab
Acquired : 01 Jul 2021 15:25 using AcqMethod CHICK_2019.M
Instrument : GCMS 6
Sample Name: CP-MD-03
Misc Info : Maryland Cannabis Pesticide Mix in AccN
Vial Number: 112



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C



Not ID'd by GC/MS - Abamectin (71751-41-2), Ethephon (16672-87-0)