


File : \\SVR-FS-01\Organic QC\QC Data\GCMS-06 Minimal\DATA\082522\2
 ... 1- KM.D
 Operator : Organic QC Lab
 Instrument : GCMS 6
 Acquired : 25 Aug 2022 17:59 using AcqMethod CHICK_2019.M
 Sample Name: CP-PA-01-R1
 Misc Info : Pennsylvania Cannabis Standard in AcCN



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
 Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
 GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
 Split 20:1, 1µl inj.; GC/MS; Injector 270°C



Not ID'd by GC/MS - Abamectin (71751-41-2), Acequinocyl (57960-19-7), Aldicarb (116-06-3)