

File :K:\QC Data\GCMS-04 Minimal\DATA\040620\01-.D
Operator : RFXM
Acquired : 7 Apr 2020 8:27 using AcqMethod VOC-100SPLIT DECKER.M
Instrument : GCMS4
Sample Name: CP-TER-011S
Misc Info : RT
Vial Number: 1



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

