



=====

Acq. Operator : SYSTEM Seq. Line : 19
 Acq. Instrument : LCMS-1 Location : P1-C4
 Injection Date : 11/22/2022 12:23:51 PM Inj : 1
 Inj Volume : 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
 Last changed : 11/5/2021 2:37:17 PM by SYSTEM
 Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
 Last changed : 11/9/2022 8:35:37 AM by SYSTEM
 (modified after loading)

Method Info : Operator: Organic QC lab
 Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

