

File :Q:\GCMS-04\DATA\030519\23-.D  
Operator : RFXM  
Acquired : 6 Mar 2019 1:51 using AcqMethod VOC-100SPLIT DECKER.M  
Instrument : GCMS4  
Sample Name: D-3606-IS3-05  
Misc Info :  
Vial Number: 66



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm  
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)  
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

