

File :D:\MassHunter\GCMS\1\DATA\060923\15 - SAK.D
Operator : Organic QC Lab
Acquired : 09 Jun 2023 09:05 using AcqMethod CHICK_2019_VOC-SV_5min.M
Instrument : GCMS 6
Sample Name : D-4629-91-LB-1X
Misc Info : 1.0 ug/mL
Vial Number: 19



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 35°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

