

File :Q:\GCMS-07 Minimal\DATA\08052002\13- -DC.D
Operator : Organic QC department
Acquired : 05 Aug 2020 17:54 using AcqMethod KERLA-2017_Short_15Min_S100.M
Instrument : 5977B MSD
Sample Name: D-4629-91-LB-BL
Misc Info : Isooctane Blank
Vial Number: 107



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 24c/min to 340c, 2min
GC Parameters: Cons. Split, 5.4psi constant flow
Split 100:1 1uL inj.; GC/MS; INJ 270c

