

File :M:\GCMS-08 Minimal\DATA\20220204\14- ALR.D  
Operator : Organic QC Lab  
Acquired : 04 Feb 2022 12:09 using AcqMethod Hy\_Hydrocarbons.M  
Instrument : GCMS 8  
Sample Name: D-5442-RT1  
Misc Info : ASTM D5442 Retention Time Standard Mix 1 in Heptane  
Vial Number: 20



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

