

=====  
Acq. Operator : SYSTEM Seq. Line : 10  
Acq. Instrument : LCMS-1 Location : P2-F4  
Injection Date : 6/17/2022 10:02:37 AM Inj : 1  
Inj Volume : 1.000 µl

Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 11/5/2021 2:37:17 PM by SYSTEM  
Method Info : Operator: Organic QC lab  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %  
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate  
:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron  
particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadruple  
MS Detector. No UV Reference.

Sample Info : 1-(2-Pyridyl)piperazine derivative of 2,4-TDI in DMSO

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

