

File :D:\MassHunter\GCMS\1\DATA\041922\20- ALR.D  
Operator : Organic QC Lab  
Acquired : 19 Apr 2022 18:03 using AcqMethod CHICK\_2019\_2S\_2uL.M  
Instrument : GCMS 6  
Sample Name: D-701S  
Misc Info : RT: 1,2,3,4,6,7,8-Heptachlorodibenzo-p-dioxin in Toluene  
Vial Number: 4



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

