Print of window 38: Current Chromatogram(s)
Data File : D:\DATA\09171901\09171901 2019-09-18 09-29-52\18- .D Sample Name : DIN38407-17

Acq. Operator : SYSTEM Seq. Line : 4
Acq. Instrument : LCMS-1 Location : P1-B9
Injection Date : 9/18/2019 10:26:48 AM Inj : 1
Inj Volume : 1.000 µl
Different Inj Volume from Sample Entry! Actual Inj Volume : 0.200 µl
Acq. Method : D:\DATA\09171901\09171901 2019-09-18 09-29-52\Purity195_315.M
Last changed : 9/18/2019 9:29:52 AM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 4/10/2019 10:56:50 AM by SYSTEM
Method Info : Operator: GA
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV Reference.

Current Chromatogram(s)