

File : Q:\GCMS-08 Minimal\DATA\20220711\30- [JP].D
Operator : Organic QC Lab
Acquired : 11 Jul 22 07:03 pm using AcqMethod Hy_PAH_24cm_S50.M
Instrument : GCMS 8
Sample Name: DRH-MA-SS-10X
Misc Info in acetone
Vial Number: 117



Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

