

File :D:\MassHunter\GCMS\1\DATA\04252301\08- KD.D
 Operator : Organic QC department
 Acquired : 25 Apr 23 01:36 pm using AcqMethod KERLA-2017_S50.M
 Instrument : 5977B MSD
 Sample Name: DRH-NJ-001S
 Misc Info :
 Vial Number: 48

Column: DB-5MS, 30m, 0.25 ID, 0.25 um
 Oven Program: 80c, 17c/min to 340c, 8min
 GC Parameters: Cons. Split, 12psi constant flow
 Split 20:1, 1uL inj.; GC/MS; INJ 270c

