

File :D:\MassHunter\GCMS\1\data\11131901\50- .D  
 Operator : KD&DC  
 Acquired : 14 Nov 19 05:41 am using AcqMethod KERLA-2017\_S50.M  
 Instrument : 5977B MSD  
 Sample Name: DRO-AK-102-NAS-10X  
 Misc Info in DCM  
 Vial Number: 35

Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
 Oven Program: 80c, 17c/min to 340c, 8min  
 GC Parameters: Cons. Split, 12psi constant flow  
 Split 20:1, 1uL inj.; GC/MS; INJ 270c

