

=====
Acq. Operator : SYSTEM Seq. Line : 38
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2 Location : P1-A-08
Injection Date : 12/20/2022 12:49:27 AM Inj : 1
Inj Volume : 1.000 µl



Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\12192201\121922101 2022-12-19 13-06-13\Purity_195-315.M
Last changed : 3/23/2022 11:04:54 AM by SYSTEM
Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M
Last changed : 3/23/2022 11:04:54 AM by SYSTEM
Method Info : OPERATOR: Organic QC lab
General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Sample Info : Disperse Orange 37 in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

