

```

=====
Acq. Operator   : SYSTEM                      Seq. Line :   27
Acq. Instrument : LCMS-1                    Location  : P1-F3
Injection Date  : 2/22/2023 6:44:11 PM      Inj       :    1
                                           Inj Volume: 1.000 µl

```



```

Method          : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 1/5/2023 12:01:53 PM by SYSTEM
Method Info     : Operator: Organic QC lab
                  Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                  Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
                  /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size.
                  Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV
                  Reference.

```

Sample Info : Solvent Red 19 @ 100 ug/mL in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

