

```

=====
Acq. Operator   : SYSTEM                      Seq. Line :   10
Acq. Instrument : LCMS-1                      Location  :   P1-E9
Injection Date  : 12/21/2020 10:03:08 AM      Inj       :    1
                                                    Inj Volume: 1.000 µl
  
```

```

Acq. Method     : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 5/8/2020 11:20:44 AM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 11/30/2020 12:15:35 PM by SYSTEM
                 (modified after loading)
  
```

```

Method Info     : Operator: Oraganic QC lab
                 Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                 Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
                 /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.
                 Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV
                 Reference.
  
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

