



```

=====
Acq. Operator   : SYSTEM                               Seq. Line :   17
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                               Location  :   P2-E9
Injection Date  : 11/2/2021 10:41:16 AM                Inj       :    1
                                                    Inj Volume: 1.000 µl
    
```

```

Acq. Method     : C:\Users\Public\Documents\ChemStation\1\Data\11022101\11022101 2021-11-02 05-38-24
                  \Purity_195-315.M
Last changed    : 4/29/2021 10:41:07 AM by SYSTEM
Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M
Last changed    : 4/29/2021 10:41:07 AM by SYSTEM
Method Info     : OPERATOR: Organic QC lab
                  General gradient method starting @100% water for 1 min @0.3 ml/min,the 100%
                  Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
                  plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system
                  , Agilent 6125 quadrupole MS detector.
    
```

Current Chromatogram(s)

