

File :Q:\GCMS-06 Minimal\DATA\011821\BSB\09- KM.D
Operator : Organic QC Lab
Acquired : 18 Jan 2021 17:32 using AcqMethod CHICK_2019.M
Instrument : GCMS 6
Sample Name: FU-005
Misc Info : Kerosene in MeOH
Vial Number: 16



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

