

File : L:\QC Data\GCMS-06 Minimal\DATA\090222\11 - SAK.D  
Operator : Organic QC Lab  
Acquired : 02 Sep 2022 13:07 using AcqMethod CHICK\_2019\_VOC-SV.M  
Instrument : GCMS 6  
Sample Name: FU-009-D-10X  
Misc Info : Diesel fuel in DCM dil 10x  
Vial Number: 108

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

