

File :D:\MassHunter\GCMS\1\data\20220527\02-ALR.D
Operator : Organic QC Lab
Acquired : 27 May 2022 08:54 using AcqMethod Hy_VOC-SV_15cm_S20.M
Instrument : GCMS 8
Sample Name: FU-015)
Misc Info : RT: JP-8 Fuel in MeOH
Vial Number: 5



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

