

File :Q:\GCMS-06 Minimal\DATA\050619\17-D
Operator : D.C
Acquired : 06 May 2019 20:31 using AcqMethod CHICK_2017_S2_2uL.M
Instrument : GCMS 6
Sample Name: GPC-CC-A-WL
Misc Info : Working Level GPC Calibration in CH2Cl2
Vial Number: 61



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 2:1, 2uL inj.; GC/MS; INJ 270c

