

File :Q:\GCMS-06 Minimal\DATA\11\03-.D
Operator : KM
Acquired : 21 Nov 2019 12:49 using AcqMethod M502.M
Instrument : GCMS 6
Sample Name: GRH-004S-SS-R2
Misc Info : VPH Standard in MeOH
Vial Number: 15



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

