

File :M:\GCMS-06 Minimal\DATA\121021B\27- YX.D
Operator : Organic QC Lab
Acquired : 11 Dec 2021 03:43 using AcqMethod VOC-Split20.M
Instrument : GCMS 6
Sample Name: GRO-AK-101-SS
Misc Info :
Vial Number: 14



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

