

File :Q:\GCMS-06 Minimal\DATA\041620\15-.D
Operator : Organic QC Lab
Acquired : 16 Apr 2020 19:40 using AcqMethod CHICK_2019_2S_2uL.M
Instrument : GCMS 6
Sample Name: H-125S
Misc Info : Acenaphthylene in Toluene
Vial Number: 39



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

