

File :D:\MassHunter\GCMS\1\DATA\032522\11 -SAK.D  
Operator : Organic QC Lab  
Acquired : 25 Mar 2022 13:06 using AcqMethod CHICK\_2019\_13MIN.M  
Instrument : GCMS 6  
Sample Name: H-135S  
Misc Info : Dibenz(a,h)anthracene in Toluene  
Vial Number: 110



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 2:1, 2µl inj.; GC/MS; injector 270°C

