

File :Z:\QC Data\GCMS-07 Minimal\DATA\01152001\09-.D  
Operator : Organic QC department  
Acquired : 15 Jan 20 09:31 am using AcqMethod KERLA-2017.M  
Instrument : 5977B MSD  
Sample Name: H-138S  
Misc Info in Toluene  
Vial Number: 71



Column: DB-5MS, 30m, 0.25 ID, 0.25  $\mu$ m  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

