

File :Q:\GCMS-08 Minimal\DATA\20210525\02- YX.D
Operator : Organic QC Lab
Acquired : 25 May 2021 09:57 using AcqMethod Hy_PAH_24cm_S1.M
Instrument : GCMS 8
Sample Name: H-149S
Misc Info : 9-Methylanthracene in toluene
Vial Number: 111



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

