

File :D:\MassHunter\GCMS\1\data\20220427\12 - SAK.D
Operator : Organic QC Lab
Acquired : 27 Apr 2022 11:30 using AcqMethod Hy_Pest_24cm_S10.M
Instrument : GCMS 8
Sample Name: H-201S
Misc Info :
Vial Number: 15



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

