

File : Q:\GCMS-06 Minimal\DATA\020821\25- -DC.D
Operator : Organic QC Lab
Acquired : 08 Feb 2021 21:22 using AcqMethod CHICK_2019.M
Instrument : GCMS 6
Sample Name : H-217S
Misc Info : 5-Methylbenz(a)anthracene in Toluene
Vial Number: 30



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

