

File :D:\MassHunter\GCMS\1\data\09072022\11-ALR.D

Operator :
Acquired : 07 Sep 2022 11:57 using AcqMethod Hy SV DB-5MS UI20M 0.18 40 split TIME17 MIN .M
Instrument : GCMS 9
Sample Name: H-242S
Misc Info : Cyclopenta(c,d)pyrenein Toluene
Vial Number: 7



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

