

File :L:\QC Data\GCMS-08 Minimal\DATA\20220526\64 - SAK.D  
Operator : Organic QC Lab  
Acquired : 27 May 2022 07:56 using AcqMethod Hy\_VOC-SV\_15cm\_S20.M  
Instrument : GCMS 8  
Sample Name: HS-002S-40X  
Misc Info :  
Vial Number: 6



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

