

File :D:\MassHunter\GCMS\1\data\053019\33-.D  
Operator : D.C  
Acquired : 31 May 19 04:35 am using AcqMethod CHICK\_2017\_S2\_2uL.M  
Instrument : GCMS 6  
Sample Name: ISO6468-PCB  
Misc Info : ISO 6468 PCB Std in Hexane  
Vial Number: 46



Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 2:1, 2uL inj.; GC/MS; INJ 270c

