



=====

Acq. Operator : SYSTEM Seq. Line : 2

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2 Location : P2-D-01

Injection Date : 10/20/2022 6:53:30 AM Inj : 2

Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\10202201\10202201 2022-10-20 06-35-41\DNPH.M

Last changed : 12/4/2019 12:27:07 PM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M

Last changed : 3/23/2022 11:04:54 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

DAD1 A, Sig=365,4 Ref=360,100 (C:\USERS\P...TA\10202201\10202201 2022-10-20 06-35-41\02-KD.D)

