

File : Q:\GCMS-08 Minimal\DATA\20190424\BSB\14- D
Operator : JP/KM
Acquired : 24 Apr 2019 20:51 using AcqMethod BDE209_LONG.M
Instrument : GCMS 8
Sample Name: M-1614-CM
Misc Info : PBDE Congeners in Isooctane
Vial Number: 90



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

