

File :D:\MassHunter\GCMS\1\DATA\08202101\19- ALR.D  
Operator : Organic QC department  
Acquired : 20 Aug 2021 12:52 using AcqMethod KERLA-2019\_Short\_15Min\_2S\_2uL.M  
Instrument : 5977B MSD  
Sample Name: M-1618-SP  
Misc Info : RT: Tributylphosphate & Triphenyl phosphate in Acetone  
Vial Number: 58



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

