

File :D:\MassHunter\GCMS\1\data\10031901\03- .D
Operator : KD
Acquired : 03 Oct 19 09:28 am using AcqMethod KERLA-2017_Short_15Min.M
Instrument : 5977B MSD
Sample Name: M-1626-01S
Misc Info :
Vial Number: 31

Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

