

File :Q:\GCMS-08 Minimal\DATA\20200609\27- -DC.D
Operator : Organic QC Lab
Acquired : 09 Jun 20 06:35 pm using AcqMethod Hy_Pest_15cm_S10.M
Instrument : GCMS 8
Sample Name: M-1653-TS
Misc Info : DFTPP in Acetone
Vial Number: 126



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 24c/min to 340c, 2min
GC Parameters: Cons. Split, 5.4psi constant flow
Split 10:1, 1uL inj.; GC/MS; INJ 270c

