

File : Q:\GCMS-06\DATA\021519-2\BSB\02-D
Operator : D.C
Acquired : 15 Feb 2019 15:20 using AcqMethod CHICK_2017_Splitless_Plus 400EMV.M
Instrument : GCMS 6
Sample Name: M-1657-02-1X
Misc Info : Calibration Group 2 in Isooctane
Vial Number: 49



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

