

```
=====
Acq. Operator   : SYSTEM                               Seq. Line :    5
Acq. Instrument : LCMS-1                             Location  : P1-C1
Injection Date  : 1/9/2020 2:42:23 PM                Inj       :    1
                                                    Inj Volume: 1.000 µl
```

```
Acq. Method    : C:\CHEM32\1\METHODS\DNPH.M
Last changed   : 4/2/2018 12:14:16 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 10/18/2019 4:18:14 PM by SYSTEM
Method Info    : Operator: GA
                Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
                /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.
                Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV
                Reference.
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

