

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-08 Minimal\DATA\20211213
... \45- KM.D
Operator : Organic QC Lab
Instrument : GCMS 8
Acquired : 14 Dec 2021 00:58 using AcqMethod Hy_Pest_15cm_S100.M
Sample Name: M-1673-DERV-5ML
Misc Info : 3,5-Dinitrobenzyl chloride in THF



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

