

File : Q:\GCMS-06 Minimal\DATA\052322\41- RM.D
Operator : Organic QC Lab
Acquired : 24 May 2022 13:26 using AcqMethod VOC20Min#.M
Instrument : GCMS 6
Sample Name: M-502-32
Misc Info :
Vial Number: 102

Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

