

File : B:\QC Data\GCMS-04 Minimal\DATA\100319\27-D
Operator : RM / KM
Acquired : 4 Oct 2019 00:25 using AcqMethod VOC-100SPLIT DECKER.M
Instrument : GCMS4
Sample Name: M-502-45-10X
Misc Info :
Vial Number: 14



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

