

File :Q:\GCMS-04 Minimal\DATA\101221\11- YX.D
 Operator : Organic QC Lab
 Acquired : 12 Oct 2021 14:08 using AcqMethod VOC-50SPLIT DECKER.M
 Instrument : GCMS4
 Sample Name: M-502
 Misc Info : Method 502.2 - Volatiles Organic
 Vial Number: 6



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
 Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
 GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

