

File : D:\MassHunter\GCMS\1\DATA\01062001\28- .D
 Operator : Organic QC department
 Acquired : 06 Jan 20 10:20 pm using AcqMethod KERLA-2017.M
 Instrument : 5977B MSD
 Sample Name: M-507G
 Misc Info : in MtBE
 Vial Number: 64

Column: DB-5MS, 30m, 0.25 ID, 0.25 um
 Oven Program: 80c, 17c/min to 340c, 8min
 GC Parameters: Cons. Split, 12psi constant flow
 Split 20:1, 1uL inj.; GC/MS; INJ 270c

