

File : L:\QC Data\GCMS-08 Minimal\DATA\20220831\02 SAK.D
Operator : Organic QC Lab
Acquired : 31 Aug 2022 08:10 using AcqMethod Hy_VOC-SV_24cm_S100.M
Instrument : GCMS 8
Sample Name: M-524-IS-2-10X
Misc Info : Fluorobenzene dil 10x
Vial Number: 137



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

