

File :D:\MassHunter\GCMS\1\data\20230608\26- =JP=.D
Operator : Organic QC Lab
Acquired : 08 Jun 2023 13:06 using AcqMethod Hy_PAH_24cm_S50.M
Instrument : GCMS 8
Sample Name: M-525-FS-1
Misc Info : in acetone
Vial Number: 33



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

